

इंटरनेट

मानक

Disclosure to Promote the Right To Information

Whereas the Parliament of India has set out to provide a practical regime of right to information for citizens to secure access to information under the control of public authorities, in order to promote transparency and accountability in the working of every public authority, and whereas the attached publication of the Bureau of Indian Standards is of particular interest to the public, particularly disadvantaged communities and those engaged in the pursuit of education and knowledge, the attached public safety standard is made available to promote the timely dissemination of this information in an accurate manner to the public.

“जानने का अधिकार, जीने का अधिकार”

Mazdoor Kisan Shakti Sangathan

“The Right to Information, The Right to Live”

“पुराने को छोड़ नये के तरफ”

Jawaharlal Nehru

“Step Out From the Old to the New”

IS 12344 (1988): Reduction Forceps [MHD 17: Health Informatics]



“ज्ञान से एक नये भारत का निर्माण”

Satyanarayan Gangaram Pitroda

“Invent a New India Using Knowledge”



“ज्ञान एक ऐसा खजाना है जो कभी चुराया नहीं जा सकता है”

Bhartrhari—Nitiśatakam

“Knowledge is such a treasure which cannot be stolen”

BLANK PAGE




Indian Standard

(Reaffirmed 2006)

SPECIFICATION FOR REDUCTION FORCEPS

- 1. Scope** — This standard covers requirements for reduction forceps used in veterinary surgery.
- 2. Shape and Dimensions** — Shall conform to Fig. 1.



All dimensions in millimetres.

FIG. 1 REDUCTION FORCEPS

Adopted 25 February 1988

© October 1988, BIS

Gr 1

2.1 The tolerance on dimensions of working ends shall conform to 'fine' class of deviation and the tolerance for remaining dimensions shall conform to 'medium' class of deviation of IS : 2102 (Part 1)-1980 'General tolerances for dimensions and form and position: Part 1 General tolerances for linear and angular dimensions (*second revision*)'.

3. Materials

3.1 *Arms* — Stainless steel conforming to designation 30 Cr 13 of IS : 6603-1972 'Specification for stainless steel bars and flats'.

3.2 *Screws* — Stainless steel conforming to designation 20 Cr 13 of IS : 6603-1972.

4. Requirements

4.1 The instrument shall be free from residual scales, acids, grease and grinding and polishing materials. Compliance with these requirements shall be checked under 2× magnification.

4.2 The surface finish shall be mirror polished. The instrument should be adequately ground to remove all surface imperfections and polished to remove grinding marks, resulting in a mirror finish. The mirror finish should be effected by an appropriate procedure, such as polishing, brushing, electropolishing and mirror buffing.

4.3 The forceps shall be free from burrs, pits, crevices and cracks, when examined under 10× magnification.

4.4 *Passivation and Final Treatment* — The instruments shall be treated by a suitable passivation process.

Note — Examples of methods of passivation are by electropolishing or by treating with 10 percent (v/v) nitric acid solution for not less than 30 min at a temperature of not less than 10°C and not exceeding 60°C. The instruments should then be rinsed in water and dried in hot air.

4.5 The teeth on the jaws shall be well formed and clean.

4.6 The two arms shall move freely and smoothly at the joint without any undue play.

5. *Hardness* — The forceps shall be uniformly hardened and tempered to give a hardness of 380 to 420 HV.

6. Tests

6.1 *Corrosion Resistance Test* — The forceps shall be tested in accordance with IS : 7531-1975 'Method for boiling and autoclaving test for corrosion resistance of stainless steel surgical instruments'.

6.2 *Performance Test* — A piece of hardwood of minimum thickness 12 mm shall be held in the jaws of the forceps. A compressive force of 1 000 N (100 kgf approx) shall be applied to the forceps at the point of maximum convexity. The force should be allowed to act for two minutes. On completion of the test, the forceps shall not show any sign of damage or permanent set.

7. *Marking* — The forceps shall be marked with the manufacturer's name, initials or recognized trade-mark.

7.1 *Standard Mark* — Details available with the Bureau of Indian Standards.

8. *Packing* — As agreed to between the purchaser and the supplier.

9. *Sampling* — Unless otherwise agreed to between the purchaser and the supplier, the sampling procedure given in IS : 2500 (Part 1)-1973 'Sampling inspection tables: Part 1 Inspection by attributes and by count of defects', shall be followed.

9.1 *Lot* — All the forceps of same size, shape and belonging to the same batch of manufacture shall be grouped together to constitute a lot.

9.2 For dimensions and visual examination given in 2 and 4, a single sampling plan with Inspection Level IV and AQL 4 percent given in Table 1 and 2 of IS : 2500 (Part 1)-1973 shall be followed.

9.3 For requirements given in 5 and 6, a single sampling plan with Inspection Level I and AQL 1 percent given in Table 1 and 2 of IS : 2500 (Part 1)-1973 shall be followed.